



ORDEN DE MANTENIMIENTO CONCENTRADORES

FORMATO
Codigo: SAC-QM-FR-028
Version: 1
Fecha Actualización: 02/08/2015
Página: 1 de 2

Fecha Recepción: 10-22, Revisión #: 169366, Auxiliar Mantenimiento: M. Patino, No. Orden: [blank]

Identificación del equipo: Código Equipo: NN-3388B, Fecha Inicio: 3-10-22, Fecha Fin: 10-10-22

Equipo contaminado por posible plaga: SI [] NO [] Fecha Inicio Aislamiento: [] Fecha Fin Aislamiento: []

Table with columns: FALLA, DAÑO, CLASE DAÑO, DESCRIPCIÓN, CANTIDAD, Q. REPORTE, REPARO CAMBIADO. Includes a 'DETALLES ADICIONALES DADOS POR PACIENTE' section.

Table for 'PRUEBAS DE ALARMA' with columns for various tests (Desconexión, Fuga, Flujo, etc.) and 'OBSERVACIONES' containing handwritten notes like 'MHO' and 'compresor, cambio de mangueras, 07/01/22'.

Table for 'PRUEBA' with columns: No., Prueba, Descripción, Resultado. Includes tests for battery voltage, air leak, electrical connection, and alarm.

Table for 'CONSUMO' with columns: CANT, DESCRIPCIÓN, CANT, DESCRIPCIÓN, CANT. Includes handwritten entry: 'Cambio mangueras kit Anillo'.

Table for 'Prueba / Simulación' with columns: No., Prueba / Simulación, Descripción, Resultado. Includes tests for alarm activation and sensor response.

Table for 'PRUEBAS DE ALARMA' with columns for various tests and 'OBSERVACIONES' containing handwritten notes like 'Aprobado' and 'Naniela Patino'.