



ORDEN DE MANTENIMIENTO CONCENTRADORES

FORMATO

Coligo: SAC-MMA-PM-020
Version: 1 Fecha Actualización: 26/06/2018
Página: 1 de 3

Fecha Remisión: 21-9-21 Revisión: 1 68524 Auxiliar Mantenimiento: M. Botino No. Orden:

Identificación del Equipo: Código Equipo: AN-2003B Fecha Inicio: 21-9-21 Fecha Fin:

Equipo contaminado por posible plaga: SI NO FECHA INICIO AISLAMIENTO FECHA FIN AISLAMIENTO

Table with columns: FALLA, DÑO, CLASE DÑO, DESCRIPCIÓN, CANTIDAD, SI REPORTE, REPUESTO CAMBIADO. Includes a section for 'DETALLE DE ADICIONALES DADOS POR PACIENTE'.

Form with sections: 1. CICLADO, 2. RUIDO EXTRAÑO, 5. DAÑOS IDENTIFICADOS, 6. ESTADO DE LAS CONDICIONES, 7. CONTROL DE TEMPERATURA. Includes handwritten notes like 'se cambian mangueras' and 'BUENAS'.

Table with columns: No., Prueba, Descripción, Resultado. Contains diagnostic test results for voltage, leaks, electrical connections, and alarm.

Table with columns: CODIGO, DESCRIPCIÓN, CANT. Includes handwritten entry: 'kd de mangueras 4'.

Table with columns: No., Prueba / Simulación, Descripción, Resultado. Contains simulation test results for alarm activation and valve operation.

Form with columns: HORÓMETRO, PRESIÓN, OXÍGENO, FLOJOMETRO, PRUEBAS DE ALARMA, REJILLAS, FILTROS, BARRAS, RUEDAS. Includes handwritten notes and signature 'M. Botino'.